

# **Speed Measurement of Surface Propagating Acoustic Wave by Full-Field Laser-Ultrasound Imaging**

**Taeil Yoon<sup>1</sup>, Younggwe Kim<sup>1</sup>, Muhammad Awais<sup>1</sup>, Byeong Ha Lee<sup>1</sup>**

<sup>1</sup>School of Electrical Engineering and Computer Science, Gwangju Institute of Science and Technology,  
Republic of Korea

Glaucoma is a prevalent condition affecting 5.1% of adults aged 40 and above, and the number of glaucoma patients is significantly increasing each year. The early diagnosis and timely treatment are crucial because delayed detection can lead to optic nerve damage and potential vision loss. Measuring intraocular pressure is essential for diagnosing glaucoma. Conventional contact-based methods of measuring intraocular pressure require the inconvenience of administering eye drops for an anesthesia and carry the risk of infection or injury. To solve these problems, the non-contact tonometry using an air-puff technique has been introduced, but it can cause discomfort for patients. Additionally, inaccurate measurements may result if the physical characteristics of the actual eye are not properly considered. To overcome the limitations of the non-contact method, we have developed the full-field laser ultrasound (FF-LUS) imaging technique that can precisely measure the speed of acoustic wave propagating along the surface of an eye, enabling the measurement of intraocular pressure. In this study, the two-dimensional surface displacement caused by the generated acoustic wave through the photoacoustic effect is obtained with a home-made FF-LUS system. By using the Fourier-Bessel transform, the high spatial frequency noise is considerably reduced, which enhances the accuracy in measuring the velocity of the surface wave.